

# COMPLIANCE CLASS



## Scott Specialty Gases

### Dual-Analyzed Calibration Standard

1290 COMBERMERE STREET, TROY, MI 48083

Phone: 248-589-2950

Fax: 248-589-2134

## CERTIFICATE OF ACCURACY: EPA Protocol Gas

### Assay Laboratory

SCOTT SPECIALTY GASES  
1290 COMBERMERE STREET  
TROY, MI 48083

P.O. No.: 54942-71-65800  
Project No.: 05-43866-002

### Customer

CLEAN AIR INSTRUMENT RENTAL  
GARY ZAPPEL  
500 WEST WOOD STREET  
PALATINE IL 60067

### ANALYTICAL INFORMATION

This certification was performed according to EPA Traceability Protocol For Assay & Certification of Gaseous Calibration Standards; Procedure G-1; September, 1997.

**Cylinder Number:** ALM038890    **Certification Date:** 15 Jun 2006    **Exp. Date:** 14 Jun 2009  
**Cylinder Pressure\*\*\*:** 1500 PSIG

### ANALYTICAL

#### ACCURACY\*\*

#### TRACEABILITY

#### COMPONENT

#### CERTIFIED CONCENTRATION (Moles)

METHANE	844.0	PPM	+/- 2%	GMIS
NITROGEN		BALANCE		

\*\*\* Do not use when cylinder pressure is below 150 psig.

\*\* Analytical accuracy is based on the requirements of EPA Protocol procedures, September 1997.

### REFERENCE STANDARD

<u>TYPE/SRM NO.</u>	<u>EXPIRATION DATE</u>	<u>CYLINDER NUMBER</u>	<u>CONCENTRATION</u>	<u>COMPONENT</u>
GMIS CH4/N2	03Apr2008	K001886	100.5 PPM	METHANE

### INSTRUMENTATION

<u>INSTRUMENT/MODEL/SERIAL#</u>	<u>DATE LAST CALIBRATED</u>	<u>ANALYTICAL PRINCIPLE</u>
BECKMAN/400 A/2000630	05Jun2006	FLAME IONIZATION

APPROVED BY: \_\_\_\_\_